

Certificate of Analysis



MagMAX CORE Nucleic Acid Purification Kit

Product No. **A32702**
Lot No. **1808018**
Date of Manufacture **29AUG2018**
Expiration Date **27AUG2019**

TEST	SPECIFICATION	RESULT
PROTEASE K		
Rnase		Pass
Exonuclease		Pass
Endonuclease		Pass
Average unit activity 0.5- 1.0 U/μL		Pass
LYSIS SOLUTION		
pH	7.25 - 7.45 -	Pass
Conductivity	19 - 25 mS/cm	Pass
WASH SOLUTION 1		
Density	0.98 - 1.01 g/mL	Pass
MAGNETIC BEADS		
Rnase		Pass

For Veterinary Use Only.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Austin, TX, USA.

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

**Quality Assurance
Issued 16OCT18**

Certificate of Analysis

ThermoFisher
S C I E N T I F I C

MagMAX CORE Nucleic Acid Purification Kit

Product No. **A32702**
Lot No. **1808018**
Date of Manufacture **29AUG2018**
Expiration Date **27AUG2019**

TEST	SPECIFICATION	RESULT
ELUTION SOLUTION		
Rnase		Pass
Exonuclease		Pass
Endonuclease		Pass
pH	6.14 - 8.40 -	Pass
WASH SOLUTION 2		
Rnase		Pass
Exonuclease		Pass
Endonuclease		Pass
Density	0.82 - 0.88 g/mL	Pass
pH	6.79 - 7.67 -	Pass

For Veterinary Use Only.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Austin, TX, USA.

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

**Quality Assurance
Issued 16OCT18**