

# Certificate of Analysis

## MagMAX CORE Nucleic Acid Purification Kit

Product No.                    **A32702**  
Lot No.                         **1810022**  
Date of Manufacture       **23OCT2018**  
Expiration Date             **24SEP2019**

TEST	SPECIFICATION	RESULT
<b>PROTEASE K</b>		
<= 30% change in probe relative to control		
Rnase		Pass
Exonuclease		Pass
Endonuclease		Pass
Average unit activity	0.5- 1.0 U/μL	Pass
<b>LYSIS SOLUTION</b>		
pH	7.25 - 7.45	Pass
Conductivity	19 - 25 mS/cm	Pass
<b>WASH SOLUTION 1</b>		
Density	0.98 - 1.01 g/mL	Pass
<b>MAGNETIC BEADS</b>		
<= 30% change in probe relative to control		

*For Veterinary Use Only.*

**Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Austin, TX, USA.**

Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

**Quality Assurance**  
**Issued 14JAN19**

# Certificate of Analysis

## MagMAX CORE Nucleic Acid Purification Kit

Product No.                    **A32702**  
Lot No.                         **1810022**  
Date of Manufacture         **23OCT2018**  
Expiration Date               **24SEP2019**

TEST	SPECIFICATION	RESULT
Rnase		Pass

### ELUTION SOLUTION

<= 30% change in probe relative to control

Rnase		Pass
Exonuclease		Pass
Endonuclease		Pass
pH	6.14 - 8.40	Pass

### WASH SOLUTION 2

<= 30% change in probe relative to control

Rnase		Pass
Exonuclease		Pass
Endonuclease		Pass
Density	0.82 - 0.88 g/mL	Pass

***For Veterinary Use Only.***

***Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Austin, TX, USA.***

Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

**Quality Assurance**  
**Issued 14JAN19**

# Certificate of Analysis



## MagMAX CORE Nucleic Acid Purification Kit

Product No.                    **A32702**  
Lot No.                         **1810022**  
Date of Manufacture        **23OCT2018**  
Expiration Date              **24SEP2019**

TEST	SPECIFICATION	RESULT
pH	6.79 - 7.67	Pass

*For Veterinary Use Only.*

*Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Austin, TX, USA.*

Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

**Quality Assurance**  
**Issued 14JAN19**