

Certificate of Conformance

This certifies that the product detailed below has been processed according to our Quality System (ISO 9001:2015) and conforms to the agreed specifications in effect at the time of manufacture.

PRODUCT CODE : LBS60001X
GRADUATION : N/A
SODIUM THIOSULPHATE : N/A
GAMMA STERILISED : GM027180 (Refer to attached sheet)
BATCH NO. : H1142
EXPIRY DATE : 03/2023
MANUFACTURER : Labserv Division of Thermo Fisher Scientific NZ Ltd
ORIGIN : New Zealand
DATE : 31/03/2021

Conformed by:



Flordeliza Angeles
QA Technician

LS-032
Issue 06/0718

Product Code: GM027180
Part: 000108 (1/1/20)

CUSTOMER COPY



MSD

Animal Health

Certificate of Irradiation

Certificate No: GM027180

Customer Details

Customer: LBS60001X/R - ~~H1148~~ / ~~H1248~~ ^{H1142 H1247} *PH* 31/03/21
Customer Order No: P20736
Carton Quantity: 163

Irradiation Details

Date of Unloading: 26 March 2021
Start Label: 2086361
Finish Label: 2086523
Specified Dose Range: 15 - 32 kGy
Lowest Recorded Dose: 16 kGy
Highest Recorded Dose: 24 kGy

Product Release

This document certifies that the product as detailed above has been irradiated within the dose range as specified by the customer.
Dose results are obtained as per SOP/GM15/001 & SOP/GM16/001

Gamma Approval:

Name: A. Rennie

Sign:

Date:

29 MAR 2021

Head Licensee:

Name: D. Harris

Sign:

Date:

30 MAR 2021

TR 1279802