

Certificate of Conformance

This certifies that the product detailed below has been processed according to our Quality System (ISO 9001:2015) and conforms to the agreed specifications in effect at the time of manufacture.

PRODUCT CODE : LBSCT5203X
GRADUATION : N/A
SODIUM THIOSULPHATE : N/A
GAMMA STERILISED : GM027440 (Refer to attached sheet)
BATCH NO. : H1850
EXPIRY DATE : 06/2023
MANUFACTURER : Labserv Division of Thermo Fisher Scientific NZ Ltd
ORIGIN : New Zealand
DATE : 21/07/2021

Conformed by:



Flordeliza Angeles
QA Technician

LS-032
Issue 06/0718

5 Allens Road
East Tamaki
Auckland 2013
New Zealand

PO Box 51481
Pakuranga
Auckland 2140
New Zealand

+64 (09) 265 3320
+64 (09) 274 4174 fax

www.thermofisher.co.nz



MSD

Animal Health

Certificate of Irradiation

Certificate No: GM027440

Customer Details

Customer: **LB SCT5203X - H1850**
Customer Order No: **P21004**
Carton Quantity: **2**

Irradiation Details

Date of Unloading: **14 July 2021**
Start Label: **2103859**
Finish Label: **2103860**
Specified Dose Range: **15 - 17 kGy**
Lowest Recorded Dose: **16 kGy**
Highest Recorded Dose: **16 kGy**

Product Release

This document certifies that the product as detailed above has been irradiated within the dose range as specified by the customer. Dose results are obtained as per SOP/GM15/001 & SOP/GM16/001

Gamma Approval:

Name: L. Gallagher

Sign:

Date: 16 July 21

Head Licensee:

Name: D. Harris

Sign:

Date: 16 JUL 2021

TR 1379692