

Certificate of Conformance

This certifies that the product detailed below has been processed according to our Quality System (ISO 9001:2015) and conforms to the agreed specifications in effect at the time of manufacture.

PRODUCT CODE : LBSPD1002X (Dish Petri 90mm PS Sterile)
GRADUATION : N/A
SODIUM THIOSULPHATE : N/A
GAMMA STERILISED : 153323/ 153324/ 153749/ 153809/ 158051
STERILITY DOSE : Gamma Irradiated at 10 kilograys
BATCH NO. : S6603
EXPIRY DATE : 05/2023
MANUFACTURER : Labserv Division of Thermo Fisher Scientific NZ Ltd
ORIGIN : New Zealand
DATE : 19/12/2022

Conformed by:



Flordeliza Angeles
QA Technician

<p>LOT NO : 153323</p> <p>DATE : 22/06/2021</p> <p>CUSTOMER NUMBER : VA1130</p> <p>CUSTOMER NAME : THERMO FISHER SCIENTIFIC NEW ZEALAND LTD</p> <p>PURCHASE ORDERS : PO00020880</p> <p>OTHER REFERENCE : S6553, S6579, S6609, S6603, S6602</p>	<p style="text-align: center;">STERITECH PTY. LTD. CERTIFICATE OF IRRADIATION</p> <p>160 SOUTH GIPPSLAND HWY DANDENONG VIC. 3175</p> <p>Telephone: (03) 8726-5666 Fax: (03) 9701-3158</p>	<p style="text-align: center;">CUSTOMER COPY</p> <p>Released By: <u>QA / Manager</u></p> <p>Dose Meets Requirements: <u>Yes / No</u> <u>12-1</u></p> <p>Process Complete: <u>Yes / No</u> <u>12-2</u></p> <p>Released: Qty Signed <u>87</u></p> <p>Date: <u>1 JUL 2021</u></p> <p>Date: _____</p> <p>Date: _____</p> <p>Despatched: Qty Signed <u>87</u></p> <p>Date: <u>1 JUL 2021</u></p> <p>Date: _____</p> <p>Date: _____</p>	<p>Routine Do:</p>																																								
<p>PALLETS: 3</p> <p>PALLET CHEP: 0</p> <p>PALLET LOSCAM: 0</p> <p>PALLET PLAIN: 3</p> <p>PALLET EXPORT: 0</p>																																											
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>PRODUCT CODE</th> <th>QTY</th> <th>LENGTH x</th> <th>WIDTH</th> <th>x</th> <th>HEIGHT</th> <th>cm</th> <th>WEIGHT</th> <th>g</th> <th>DENSITY</th> </tr> </thead> <tbody> <tr> <td>LBS3712YX</td> <td>10</td> <td>46</td> <td>40</td> <td>50</td> <td>8750</td> <td>0.10</td> <td></td> <td></td> <td></td> </tr> <tr> <td>LBSPD1002X</td> <td>48</td> <td>51</td> <td>28</td> <td>45</td> <td>8320</td> <td>0.13</td> <td></td> <td></td> <td></td> </tr> <tr> <td>LBSST120X</td> <td>29</td> <td>27</td> <td>38</td> <td>68</td> <td>5660</td> <td>0.08</td> <td></td> <td></td> <td></td> </tr> </tbody> </table>		PRODUCT CODE	QTY	LENGTH x	WIDTH	x	HEIGHT	cm	WEIGHT	g	DENSITY	LBS3712YX	10	46	40	50	8750	0.10				LBSPD1002X	48	51	28	45	8320	0.13				LBSST120X	29	27	38	68	5660	0.08					
PRODUCT CODE	QTY	LENGTH x	WIDTH	x	HEIGHT	cm	WEIGHT	g	DENSITY																																		
LBS3712YX	10	46	40	50	8750	0.10																																					
LBSPD1002X	48	51	28	45	8320	0.13																																					
LBSST120X	29	27	38	68	5660	0.08																																					
<p>LOOSE: 0</p> <p>PALLET SKID: 0</p> <p>STRETCH WRAP: Y</p> <p>SEG. BATCHES: N</p>																																											
<p>Special Comments:</p>																																											
<p>TOTAL: 87 CTNS: 6.028 M³ 651.00 kg</p> <p>REQUIRED DOSE kGy: 10.0</p>																																											

LOT NO : **153324**

DATE : **23/06/2021**

CUSTOMER NUMBER : **VA1130**

CUSTOMER NAME : **THERMO FISHER SCIENTIFIC NEW ZEALAND LTD**

PURCHASE ORDERS : **PO00020919**

OTHER REFERENCE : **S6486, S6499, S6595, S6564, S6565, S6576, S6577, S6583, S6584, S6607, S6617, S6627**

STERITECH PTY. LTD.
CERTIFICATE OF IRRADIATION
160 SOUTH GIPPSLAND HWY Telephone: (03) 8726-5566
DANDENONG VIC. 3175 Fax: (03) 9701-3158

PALLETS: **32**
PALLET CHEP: **32**

PALLET LOSCAM: **0**
PALLET PLAIN: **0**

PALLET EXPORT: **0**

PRODUCT CODE	QTY	LENGTH	WIDTH	HEIGHT	cm	WEIGHT	g	DENSITY
LBS1209WX	7	43	28	38	6280	0.14		
LBS32024W	3	40	27	43	7500	0.16		
LBS33550X	4	46	31	46	7850	0.12		
LBS33550YX	15	31	46	46	7850	0.12		
LBS60001X	163	46	31	46	8780	0.13		
LBS60014X	36	31	46	46	4000	0.06		
LBS60016X	5	38	28	43	10840	0.24		
LBS60060X	17	46	31	46	8600	0.13		
LBSCT1203X	4	43	38	28	58000	1.27		
LBSF001BX/10	9	39	28	43	5140	0.11		
LBSLP7202	24	38	28	43	7500	0.16		
LBSLP7231X/50	21	38	28	43	10840	0.24		
LBS22001X10	10	29	30	44	8500	0.22		
LBS22002X	138	51	28	45	8320	0.13		
LBS22003X	5	40	28	35	7600	0.19		
LBS2210X	5	43	38	28	7500	0.16		
LBS22001X	19	28	38	43	4760	0.10		
LBS2200X	22	30	44	38	4400	0.09		
LBS2250ENVX	10	43	38	28	1700	0.04		
LBS2250PWX	40	43	28	38	1700	0.04		
LBS2250X	20	38	27	68	3050	0.04		
LBS2400X	47	28	43	38	3000	0.07		

TOTAL: **630** CTNS: **36.524** M³ **4520.0** kg

REQUIRED DOSE **kgY:** 10.0

CUSTOMER COPY

Released By: **QA / Manager**

Dose Meets Requirements: **Yes / No**

Process Complete: **Yes / No**

Released: **Qty Signed**

Date: **25 JUN 2021 630**

Date

Despatched: **Qty Signed**

Date: **25 JUN 2021 630**

Date

Date

ROUTINE IRRADIATION
12-4, 12-6, 12-4, 12-2, 12-3,
DOSES - METHOGENA'S
12-8, 12-5, 12-3, 12-3

Special Comments:
S6565, S6577, S6603, S6602, S6634, S6546, S6523, S6623, S6562, S6452, S6518, S6544, S6547, S6570, S6550, S6549, S6529, S6635, S6545, S6594, S6552, S6630, S6467, S6528
CONTAINER bu2913148

pg 1

Product Code	QTY	Length	Width	Height	cm	Weight	g	Density	Released By	Dose Meets Requirements	Process Complete	Released	Date	Despatched	Date	Date	Date	Routine Dosimetry (kGy)
LBS1209WX	7	43	28	38	6280	0.14			QA / Manager	Yes	Yes	Qty Signed	25 JUN 2021 630					12-5
LBS32024W	3	40	27	43	7500	0.16												12-5
LBS33550X	4	46	31	46	7850	0.12												12-3
LBS33550YX	15	31	46	46	7850	0.12												12-7
LBS60001X	163	46	31	46	8780	0.13												12-5
LBS60014X	36	31	46	46	4000	0.06												12-3
LBS60016X	5	38	28	43	10840	0.24												12-1
LBS60060X	17	46	31	46	8600	0.13												12-5
LBSCT1203X	4	43	38	28	58000	1.27												12-6
LBSF001BX/10	9	39	28	43	5140	0.11												12-1
LBSLP7202	24	38	28	43	7500	0.16												12-3
LBSLP7231X/50	21	38	28	43	10840	0.24												12-1
LBS22001X10	10	29	30	44	8500	0.22												12-3
LBS22002X	138	51	28	45	8320	0.13												12-1
LBS22003X	5	40	28	35	7600	0.19												12-8
LBS2210X	5	43	38	28	7500	0.16												12-4
LBS22001X	19	28	38	43	4760	0.10												12-4
LBS2200X	22	30	44	38	4400	0.09												12-4
LBS2250ENVX	10	43	38	28	1700	0.04												12-3
LBS2250PWX	40	43	28	38	1700	0.04												12-8
LBS2250X	20	38	27	68	3050	0.04												12-4
LBS2400X	47	28	43	38	3000	0.07												12-7

<p>LOT NO : 153324</p> <p>DATE : 23/06/2021</p> <p>CUSTOMER NUMBER : VA1130</p> <p>CUSTOMER NAME : THERMO FISHER SCIENTIFIC NEW ZEALAND LTD</p> <p>PURCHASE ORDERS : PO00020919</p> <p>OTHER REFERENCE : S6486, S6499, S6595, S6564, S6565, S6576, S6577, S6583, S6584, S6607, S6617, S6627</p>	<p style="text-align: center;">STERITECH PTY. LTD.</p> <p style="text-align: center;">CERTIFICATE OF IRRADIATION</p> <p>160 SOUTH GIPPSLAND HWY DANDENONG VIC. 3175</p> <p>Telephone: (03) 8726-5566 Fax: (03) 9701-3158</p>	<p style="text-align: center;">CUSTOMER COPY</p> <p>Released By: QA / Manager</p> <p>Dose Meets Requirements: <input checked="" type="checkbox"/> Yes / <input type="checkbox"/> No</p> <p>Process Complete: <input checked="" type="checkbox"/> Yes / <input type="checkbox"/> No</p> <p>Released: Qty Signed</p> <p>Date: 25 JUN 2024 6:30 <i>ll</i></p> <p>Date: _____</p> <p>Date: _____</p> <p>Despatched: Qty Signed</p> <p>Date: 25 JUN 2024 6:30 <i>ll</i></p> <p>Date: _____</p> <p>Date: _____</p>	<p>Routine Dosimetry (kGy)</p> <p>11.5</p> <p>12.3</p> <p>12.3</p> <p>12.7</p> <p>12.5</p> <p>12.6</p> <p>12.1</p> <p>12.3</p> <p>12.1</p> <p>12.8</p> <p>12.4</p> <p>12.4</p> <p>12.6</p> <p>12.4</p> <p>12.4</p> <p>12.3</p> <p>12.8</p> <p>12.4</p> <p>12.4</p> <p>12.7</p>
<p>PALLETS: 32</p> <p>PALLET CHEP: 32</p> <p>PALLET LOSCAM: 0</p> <p>PALLET PLAIN: 0</p> <p>PALLET EXPORT: 0</p> <p>LOOSE: 0</p> <p>PALLET SKID: 0</p> <p>STRETCH WRAP: Y</p> <p>SEG. BATCHES: N</p>		<div style="border: 1px solid black; padding: 5px; margin-bottom: 10px;"> <p style="text-align: center;">ROUTINE IRRADIATION</p> <p style="text-align: center;">12.4, 12.6, 12.4, 12.2, 12.3</p> <p style="text-align: center;">DOSES - KILOGRAY'S</p> <p style="text-align: center;">12.8, 12.5, 12.3, 12.3</p> </div> <p>Special Comments:</p> <p>S6565, S6577, S6603, S6602, S6634, S6546, S6523, S6623, S6623, S6623, S6452, S6518, S6544, S6547, S6570, S6550, S6549, S6529, S6635, S6545, S6594, S6552, S6630, S6467, S6528</p> <p>CONTAINER bu2913148</p>	
<p>PRODUCT CODE QTY LENGTH x WIDTH x HEIGHT cm WEIGHT g DENSITY</p> <p>LBSUC1014X 6 43 38 28 10000 0.22</p>		<p>TOTAL: 630 CTNS: 36.524 M³ 4520.0 kg</p> <p>REQUIRED DOSE <u>10.0</u> kGy</p>	

Pg 2

<p>LOT NO : 153749</p> <p>DATE : 3/08/2021</p> <p>CUSTOMER NUMBER : VA1130</p> <p>CUSTOMER NAME : THERMO FISHER SCIENTIFIC NEW ZEALAND LTD</p> <p>PURCHASE ORDERS : P000021007</p> <p>OTHER REFERENCE : 86628, 86602, 86603, 86611, 86633, 86646, 86651, 86663, 86668, 86675, 86678</p>	<p>STERITECH PTY. LTD. CERTIFICATE OF IRRADIATION</p> <p>160 SOUTH GIPPSLAND HWY DANDENONG VIC. 3175</p> <p>Telephone: (03) 8728-5688 Fax: (03) 8701-3158</p>	<p>CUSTOMER COPY</p> <p>Released By: QA / Manager</p> <p>Dose Meets Requirements: <input checked="" type="checkbox"/> Yes / <input type="checkbox"/> No</p> <p>Process Complete: <input checked="" type="checkbox"/> Yes / <input type="checkbox"/> No</p> <p>Released: _____ City _____</p> <p>Date: <u>6 Aug 2021</u> <u>370</u></p> <p>Date: _____</p> <p>Despatched: _____ City _____</p> <p>Date: _____</p> <p>Date: _____</p> <p>Date: _____</p>	<p>Routine Destimetry (dGy)</p> <p>12-0</p> <p>12-1</p> <p>12-2</p> <p>12-3</p> <p>12-3</p> <p>12-2</p> <p>12-4</p> <p>12-4</p> <p>13-0</p> <p>13-0</p> <p>12-8</p> <p>12-8</p> <p>12-1</p> <p>12-5</p> <p>13-0</p> <p>12-2</p> <p>12-4</p> <p>12-4</p>																																																																																																														
<p>PALLETS: 14</p> <p>PALLET CHEP: 14</p> <p>PALLET LOSCAM: 0</p> <p>PALLET PLAIN: 0</p> <p>PALLET EXPORT: 0</p> <p>LOC-SE: 0</p> <p>PALLET SKID: 0</p> <p>STRETCH WRAP: Y</p> <p>SEG. BATCHES: N</p>																																																																																																																	
<table border="1" style="width:100%; border-collapse: collapse;"> <thead> <tr> <th>PRODUCT CODE</th> <th>QTY</th> <th>LENGTH x</th> <th>WIDTH</th> <th>x</th> <th>HEIGHT</th> <th>cm</th> <th>WEIGHT</th> <th>g</th> <th>DENSITY</th> </tr> </thead> <tbody> <tr> <td>LBSS3550YX</td> <td>18</td> <td>31</td> <td>46</td> <td>46</td> <td>7860</td> <td>0.12</td> <td></td> <td></td> <td></td> </tr> <tr> <td>LBSS3670G</td> <td>15</td> <td>28</td> <td>43</td> <td>38</td> <td>5000</td> <td>0.11</td> <td></td> <td></td> <td></td> </tr> <tr> <td>LBSS6001X</td> <td>5</td> <td>46</td> <td>31</td> <td>46</td> <td>8780</td> <td>0.13</td> <td></td> <td></td> <td></td> </tr> <tr> <td>LBSSCT5203X</td> <td>3</td> <td>28</td> <td>38</td> <td>43</td> <td>4740</td> <td>0.10</td> <td></td> <td></td> <td></td> </tr> <tr> <td>LBSP22601X</td> <td>15</td> <td>28</td> <td>43</td> <td>38</td> <td>5000</td> <td>0.11</td> <td></td> <td></td> <td></td> </tr> <tr> <td>LBSPD1002X</td> <td>74</td> <td>61</td> <td>28</td> <td>45</td> <td>8320</td> <td>0.13</td> <td></td> <td></td> <td></td> </tr> <tr> <td>LBSSPR002X</td> <td>13</td> <td>28</td> <td>43</td> <td>38</td> <td>5000</td> <td>0.11</td> <td></td> <td></td> <td></td> </tr> <tr> <td>LBSSST200X</td> <td>105</td> <td>30</td> <td>44</td> <td>38</td> <td>4400</td> <td>0.09</td> <td></td> <td></td> <td></td> </tr> <tr> <td>LBSSST250X</td> <td>120</td> <td>38</td> <td>27</td> <td>68</td> <td>3050</td> <td>0.04</td> <td></td> <td></td> <td></td> </tr> <tr> <td>LBSSST400X</td> <td>5</td> <td>28</td> <td>43</td> <td>38</td> <td>3000</td> <td>0.07</td> <td></td> <td></td> <td></td> </tr> </tbody> </table>		PRODUCT CODE	QTY	LENGTH x	WIDTH	x	HEIGHT	cm	WEIGHT	g	DENSITY	LBSS3550YX	18	31	46	46	7860	0.12				LBSS3670G	15	28	43	38	5000	0.11				LBSS6001X	5	46	31	46	8780	0.13				LBSSCT5203X	3	28	38	43	4740	0.10				LBSP22601X	15	28	43	38	5000	0.11				LBSPD1002X	74	61	28	45	8320	0.13				LBSSPR002X	13	28	43	38	5000	0.11				LBSSST200X	105	30	44	38	4400	0.09				LBSSST250X	120	38	27	68	3050	0.04				LBSSST400X	5	28	43	38	3000	0.07					
PRODUCT CODE	QTY	LENGTH x	WIDTH	x	HEIGHT	cm	WEIGHT	g	DENSITY																																																																																																								
LBSS3550YX	18	31	46	46	7860	0.12																																																																																																											
LBSS3670G	15	28	43	38	5000	0.11																																																																																																											
LBSS6001X	5	46	31	46	8780	0.13																																																																																																											
LBSSCT5203X	3	28	38	43	4740	0.10																																																																																																											
LBSP22601X	15	28	43	38	5000	0.11																																																																																																											
LBSPD1002X	74	61	28	45	8320	0.13																																																																																																											
LBSSPR002X	13	28	43	38	5000	0.11																																																																																																											
LBSSST200X	105	30	44	38	4400	0.09																																																																																																											
LBSSST250X	120	38	27	68	3050	0.04																																																																																																											
LBSSST400X	5	28	43	38	3000	0.07																																																																																																											
<p>TOTAL: 370 CTNS: 22.043 M³ 1848.1 kg</p> <p>REQUIRED DOSE <u>kgv:</u> 10.0</p>		<p>Special Comments: Container TG=U8843776.</p>																																																																																																															


<p>LOT NO : 153809</p> <p>DATE : 9/08/2021</p> <p>CUSTOMER NUMBER : VA1130</p> <p>CUSTOMER NAME : THERMO FISHER SCIENTIFIC NEW ZEALAND LTD</p> <p>PURCHASE ORDERS : PO00020992</p> <p>OTHER REFERENCE : S6546, S6575, S6600, S6602, S6603, S6623, S6649, S6651, S6663, S6666</p>	<p style="text-align: center;">STERITECH PTY. LTD. CERTIFICATE OF IRRADIATION</p> <p>160 SOUTH GIPPSLAND HWY DANDENONG VIC. 3175</p> <p>Telephone: (03) 8726-5566 Fax: (03) 9701-3158</p> <p>PALLETS: 16</p> <p>PALLET LOSCAM: 0</p> <p>PALLET PLAIN: 0</p> <p>PALLET EXPORT: 0</p> <p>LOOSE: 0</p> <p>PALLET SKID: 0</p> <p>STRETCH WRAP: Y</p> <p>SEG. BATCHES: N</p>	<p style="text-align: center;">CUSTOMER COPY</p> <p>Released By: QA / Manager</p> <p>Dose Meets Requirements: <input checked="" type="checkbox"/> Yes / <input type="checkbox"/> No</p> <p>Process Complete: <input checked="" type="checkbox"/> Yes / <input type="checkbox"/> No</p> <p>Released: Qty Signed</p> <p>Date 13 AUG 2021 313 <i>ll</i></p> <p>Date _____</p> <p>Despatched: Qty Signed</p> <p>Date 13 AUG 2021 313 <i>ll</i></p> <p>Date _____</p> <p>Date _____</p>	<p>Routine Do:</p> <p>12-1</p> <p>12-1</p> <p>12-3</p> <p>12-3</p> <p>12-3</p> <p>12-5</p> <p>12-3</p> <p>12-7</p> <p>12-5</p> <p>12-8</p> <p>12-7</p> <p>12-2</p> <p>12-7</p>
---	---	--	--

PRODUCT CODE	QTY	LENGTH	x	WIDTH	x	HEIGHT	cm	WEIGHT	g	DENSITY
LBS33036W	30	43		38		28		3000		0.07
LBS33550YX	54	31		46		46		7850		0.12
LBS60001X	68	46		31		46		8780		0.13
LBSC15203X	11	28		38		43		4740		0.10
LBSPD1002X	125	51		28		45		8320		0.13
LBSSP2210X	5	43		38		28		7500		0.16
LBSSPR002X	20	28		43		38		5000		0.11

TOTAL: 313 CTNS:	19.055 M ³	2340.6 kg
REQUIRED DOSE		10.0

Special Comments:
CONTAINER ZCSU7229792

LOT NO : 158051



DATE : 2/11/2022

CUSTOMER NUMBER : VA1130

CUSTOMER NAME : THERMO FISHER SCIENTIFIC NEW ZEALAND LTD

PURCHASE ORDERS : P000021918

OTHER REFERENCE : TLLU5199062

STERITECH PTY. LTD.
CERTIFICATE OF IRRADIATION

160 SOUTH GIPPSLAND HWY Telephone: (03) 8726-5566
DANDENONG VIC. 3175 Fax: (03) 9701-3158

OFFICE COPY

Released By:

QA / Manager

Dose Meets Requirements: Yes No

Process Complete: Yes No

Released: Qty Signed

Date 04 NOV 2022 978 *afj*

Date

Despatched: Qty Signed

Date 07 NOV 22 978 *afj*

Date

Date

PALLETS: 48

PALLET LOSCAM: 0

PALLET PLAIN: 0

PALLET EXPORT: 0

LOOSE: 0

PALLET SKID: 0

STRETCH WRAP: Y

SEG. BATCHES: N

TRANSPORT:

PRODUCT CODE	QTY	LENGTH	WIDTH	HEIGHT	cm	WEIGHT	g	DENSITY
LBS30002YX	63	40	46	50	7820	0.08		
LBS32250X	43	27	68	38	9500	0.14		
LBS32250YX	22	27	68	38	9500	0.14		
LBS32252X	15	27	38	68	9500	0.14		
LBS32252YX	47	27	68	38	9500	0.14		
LBS32260YX	15	38	66	27	5360	0.08		
LBS33550X	20	46	31	46	7850	0.12		
LBS33550YX	28	31	46	46	7850	0.12		
LBS3722W	6	46	40	50	8000	0.09		
LBS60001X	112	46	31	46	8780	0.13		
LBSLC10003X	15	50	46	40	8940	0.10		
LBSLP7201	14	28	43	38	9220	0.20		
LBSLP7202	12	38	28	43	7500	0.16		
LBSMC0003X	48	40	50	46	8940	0.10		
LBSPD1002X	48	51	28	45	8320	0.13		
LBSST250CNX	60	28	43	38	2600	0.06		
LBSST250PWX	80	43	28	38	1700	0.04		
LBSST250X	100	38	27	68	3050	0.04		
LBSST400X	102	28	43	38	3000	0.07		
LBSST500X	128	38	27	68	4900	0.07		

Special Comments:

s6981,s6992,s6993,s6695,s6947,s6994,s6758,s6923
s6953,s6696,s6964,s6857,s6955,s6948,s6828
s6983,s6919,s6950,s6984,s6978,s6988,s6963
s6789,s6976,s6987,s6603,s6657,s6962,s6934
s6933,s6935,s7008,s7009,s6997,s6943,s7003

TOTAL: 978 CTNS: 63.769 M³ 5899.6 kg

REQUIRED DOSE 10.0 kGy