



Certificate of Analysis

Capillary Array 50cm 3500xL Series

Product No. **4404689**
Lot No. **M517D79**
Expiration Date **19-Apr-2018**

4TEST	SPECIFICATION	RESULT
CAPILLARY LEAK TEST		
Each array is tested to assure absence of leaks.		
Leak Test	No Leaks	PASS
CAPILLARY PLUG TEST		
Each array is tested to assure absence of plugged capillaries.		
Plug Test	No Plugged Capillaries	PASS
FLUORESCENCE BACKGROUND TEST		
Each array is tested to assure background fluorescence is within specification.		
Fluorescence Background Test	Conforms to Specification	PASS
CAPILLARY MAPPING TEST		
Each array is tested to verify that capillary inlet and outlet spatial positioning correspond.		
Mapping Test	Conforms to Specification	PASS

For Research Use Only. Not intended for any animal or human therapeutic or diagnostic use.

**ISO 9001
REGISTERED**

Foster City, CA

Alfred Tindberg
Quality Assurance

To obtain a Certificate of Analysis on-line go to www.appliedbiosystems.com or email us at cofarequest@lifetech.com

4442861 B