

Catalog # : CHC2493


*Note: A letter at the end of the lot number signifies an additional packaging of this same lot.

1. **Coating Antibody:** **Anti-Human IL-23 Heterodimer (0.25mg/0.125 mL)**
Part Number: 58.249.09
Lot Number: 6E1/1
Form: Liquid, 1 vial, contains 0.1% sodium azide
Storage: Store at 2-8°C until expiration date.
Recommended Dilution: Dilute to 2 µg/mL with Coating Buffer B (Cat. # CB01100, or see Recommended Buffers). For example, to make 10 mL (enough to coat 1 plate), add 10 µL coating antibody to 9.990 mL Coating Buffer B.

2. **Detection Antibody:** **Anti-Human IL-23 Heterodimer Biotin (94 µg/0.125 mL)**
Part Number: 58.249.03
Lot Number: 6E1/1
Form: Liquid, 1 vial, contains 0.1% sodium azide
Storage: Store at 2-8°C until expiration date.
Recommended Dilution: Dilute to 1.5 µg/mL with Assay Buffer (Cat. # DS98200, or see Recommended Buffers). For example, to make enough for 1 plate, add 11 µL detection antibody to 5489 µL Assay Buffer.

3. **Standard:** **Recombinant Human IL-23 Heterodimer**
Part Number: 58.249.10 (additional vials of standard may be purchased using this part number)
Lot Number: 6E1/1
Form: Lyophilized, 3 vials
Storage: Store at 2-8°C.
Reconstitution: Reconstitute with Assay Buffer (Cat. # DS98200 or see Recommended Buffers) according to instructions on vial to yield a stock of 20 ng/mL. Use the standard stock immediately or aliquot into polypropylene tubes and freeze at -80°C. *Do not store at room temperature or at 4°C and do not subject to more than one freeze-thaw cycle.*
Standard Curve: Dilute standard stock to 2000 pg/mL (100 µL stock plus 900 µL Assay Buffer) with Assay Buffer (Cat. # DS98200 or see Recommended Buffers). Add 500 µL Assay Buffer to 6 tubes and label as 1000, 500, 250, 125, 62.5 and 31.25 pg/mL. Make serial dilutions starting with 2000 pg/mL by transferring 500 µL of each standard to next tube and vortexing each tube. Assay Buffer should be used as the zero standard.

4. **Streptavidin-HRP:**
Part Number: 41.000.28
Lot Number: 6E1/1
Form: Liquid, 1 vial, contains 0.05% thymol
Storage: Store at 2-8°C until expiration date.
Recommended Dilution: Dilute 1/1000 in Assay Buffer. For example, to make enough for 1 plate, add 10 µL of streptavidin-HRP to 9.990 mL of Assay Buffer (Cat. # DS98200 or see Recommended Buffers).



Representative standard curve was generated by following the recommended assay procedure, which includes the use of the **Invitrogen CytoSet™ Buffer Set** (Cat. # CNB0011)

This product is for research use only. Not for use in diagnostic procedures.

www.invitrogen.com

Invitrogen Corporation • 542 Flynn Rd • Camarillo • CA 93012 • Tel: 800.955.6288 • E-mail: techsupport@invitrogen.com

Intended Use and Materials Provided

The CytoSet™ for Human IL-23 Heterodimer contains components required to construct an enzyme-linked immunoassay for the specific and quantitative measurement of Human IL-23 Heterodimer. Sufficient quantities of all reagents are provided to yield 10 plates of 96 wells if the recommended assay procedure and recommended storage and handling of materials are followed as specified on this insert. The materials provided are **FOR RESEARCH USE ONLY**.

Recommended Buffers and Solutions

The Invitrogen CytoSet™ Buffer Set (Cat. # CNB0011) containing Coating Buffers A and B, Assay Buffer, Substrate Solution (TMB), Stop Solution, and Wash Buffer is recommended.

- Coating Buffer A:** *Coating Buffer A (Cat. # CB07100) from Invitrogen is recommended. Alternate buffer choice listed below.*
8.0 g NaCl, 1.13 g Na₂HPO₄, 0.2 g KH₂PO₄, 0.2 g KCl, 0.1% ProClin™; q.s. to 1.0 L with distilled H₂O, pH to 7.4.
- Coating Buffer B:** *Coating Buffer B (Cat. # CB01100) from Invitrogen is recommended. Alternate buffer choice listed below.*
4.3 g NaHCO₃, 5.3 g Na₂CO₃, 0.1% ProClin™; q.s. to 1.0 L with distilled H₂O, pH to 9.4.
- Assay Buffer:** *Assay Buffer (Cat. # DS98200) from Invitrogen is recommended. Alternate buffer choice listed below.*
8.0 g NaCl, 1.13 g Na₂HPO₄, 0.2 g KH₂PO₄, 0.2 g KCl, 5.0 g bovine serum albumin (fraction V), 1 mL Tween 20 and 0.5% ProClin™ as a preservative; q.s. to 1.0 L with distilled H₂O, pH to 7.4.
- Wash Buffer:** *Wash Buffer 25x (Cat. # WB01) from Invitrogen is recommended. Alternate buffer choice listed below.*
0.2 g KH₂PO₄, 1.9 g K₂HPO₄·3H₂O, 0.4 g EDTA, 0.5 mL Tween 20; q.s. to 1.0 L with distilled H₂O, pH to 7.4.
- Substrate Solution:** *TMB (Cat. # SB01) from Invitrogen is recommended. Alternate solution choice listed below.*
Tetramethylbenzidine (TMB) and Hydrogen Peroxide.
- Stop Solution:** *Stop Solution (Cat. # SS01100) from Invitrogen is recommended. Alternate solution choice listed below.*
1.8 N H₂SO₄.

Assay Optimization

CytoSets™ from Invitrogen are designed to be very flexible for your experiments. Consequently, the assay procedure contains only recommendations. The assay procedure has been optimized for use with tissue culture samples. However, serum and plasma samples may be used but may require that certain assay parameters be modified. Investigators are advised to determine optimal buffer formulations, concentrations and incubation times for individual applications.

Recommended Assay Procedure

1. Prepare coating solution by diluting the coating antibody. See “coating antibody” section for the recommended coating antibody dilution.
2. Coat plates with 100 µL per well of the coating solution. Cover plates and incubate overnight (12-18 hr.) at 4°C.
3. Aspirate wells and wash 1 time with > 400 µL of Wash Buffer per well. Following wash, invert and tap on absorbent paper to remove excess liquid.
4. Block plate with 300 µL per well of Assay Buffer for 1 hour at room temperature.
5. Aspirate, invert, and tap on absorbent paper to remove excess liquid.
6. Prepare standards and sample dilutions in Assay Buffer (or in a diluent that most closely matches the matrix of your sample). For recommended dilutions and storage of the standard, see “standard” section.
7. Pipette 200 µL of standards (in duplicate) and samples into designated wells.
8. Immediately following step 7, add 50 µL of the working detection antibody into each well. For recommended dilutions, see “detection antibody” section. **Incubate for 2 hours at 37°C.**
9. Aspirate and wash 5 times using the method in step 3.
10. Add 100 µL of the working streptavidin-HRP solution into each well. For recommended dilutions, see “streptavidin-HRP conjugate” section. **Incubate for 30 minutes at room temperature.**
11. Aspirate and wash 5 times using the method in step 3.
12. Add 100 µL of the TMB substrate to each well. Incubate plate for **30 minutes at room temperature.**
13. Add 100 µL of Stop Solution to each well.
14. Measure absorbance at 450 nm (reference absorbance: 650 nm) within 30 minutes of adding Stop Solution. Calculate results using a log-log or 4-parameter curve fit.

Additional Materials Required

- 96 well NUNC MaxiSorp microplates; NUNC Cat. # 434797.
- Pipettes, 37°C incubator and timer.
- Microplate reader with a detector that can measure absorbance at 450 nm.
- 1 L graduated cylinder; plate washer or wash bottle.
- Polypropylene tubes for standards and sample dilutions, if needed.

This product is for research use only. Not for use in diagnostic procedures.

www.invitrogen.com

Invitrogen Corporation • 542 Flynn Rd • Camarillo • CA 93012 • Tel: 800.955.6288 • E-mail: techsupport@invitrogen.com