

RapID™ ONE QUALITY CONTROL REPORT FORM

1. *Record results of biochemical reactions under the expected reactions.
 Corrective Action: See Reverse Page

Date QC performed: _____
 Technologist: _____
 ONE Plus lot #: _____
 Expiration Date: _____
 Approved by: _____

Organism	URE	ADH	ODC	LDC	TET	LIP	KSF	SBL	GUR	ONPG	BGLU	BXYL	NAG	MAL	PRO	GGT	PYR	ADON	IND	OXI
<i>P. vulgaris</i>	+	V	-	-	+	-	-	-	-	-	V	-	-	-	-	+	-	-	+	-
ATCC® 6380																				
	URE	ADH	ODC	LDC	TET	LIP	KSF	SBL	GUR	ONPG	BGLU	BXYL	NAG	MAL	PRO	GGT	PYR	ADON	IND	OXI
<i>E. coli</i> ^a	-	-	+	(+)	-	-	-	(+)	+	+	-	-	-	-	-	-	-	-	+	-
ATCC® 25922																				
	URE	ADH	ODC	LDC	TET	LIP	KSF	SBL	GUR	ONPG	BGLU	BXYL	NAG	MAL	PRO	GGT	PYR	ADON	IND	OXI
<i>P. aeruginosa</i>	V	+	V	V	V	+	-	-	-	-	-	-	-	V	+	+	V	-	-	+
ATCC® 27853																				
	URE	ADH	ODC	LDC	TET	LIP	KSF	SBL	GUR	ONPG	BGLU	BXYL	NAG	MAL	PRO	GGT	PYR	ADON	IND	OXI
<i>E. aerogenes</i> ^a	-	-	+	+	-	-	+	+	-	+	+	+	+	+	V	V	+	+	-	-
ATCC® 13048																				

+ = Positive V = Variable
 - = Negative (-) = Usually negative
 (+) = Usually positive

*Visual Interpretation Only

^aKey indicator strain

CORRECTIVE ACTION

1. List organism with biochemical reaction that is outside the acceptable criteria:

2. Date corrective action started: _____

3. Corrective action procedure performed: _____

4. Results after corrective action:

Acceptable Reaction

Outside Acceptable Reaction

5. Further corrective action results, if the biochemical reaction is still outside the acceptable criteria:

6. Date corrective results accepted: _____