	Document Owner Department: QC	BT-SPEC-0610
		Page 1 of 2
<b>OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION</b>		
<b>BACITRACIN B10 CT0005B</b>		

## Product Details

<b>Product Name</b>	BACITRACIN
<b>Disc Code</b>	B10
<b>Disc Concentration</b>	10µg
<b>CT Code</b>	CT0005B
<b>Shelf Life (years)</b>	3
<b>CLSI Recommended Concentration</b>	No
<b>EUCAST Recommended Concentration</b>	No

## Microbiological Performance Criteria


<b>Organism</b>	<b>ATCC/NCTC Number</b>	<b>Range (mm)</b>
<i>Staphylococcus aureus</i>	ATCC® 25923	11 – 17
<i>Escherichia coli</i>	ATCC® 25922	No zone
<i>Pseudomonas aeruginosa</i>	ATCC® 27853	No zone
<i>Haemophilus influenzae</i>	ATCC® 49247	No zone

## References

Tested in accordance with the current CLSI methodology, the control ranges are internally derived based on historical data.

Other additional challenging strains may be employed.

The information given above is believed to be correct, however both the information and the product are offered without warranty for any specific use, other than that specified in the current Oxoid manual.

	Document Owner Department: QC	BT-SPEC-0610
		Page 2 of 2
<b>OXOID QUALITY ASSURANCE PRODUCT SPECIFICATION</b>		
<b>BACITRACIN B10 CT0005B</b>		

### Revision History

Section / Step	Description of Change	Reason for Change	Reference
CLSI	Removal of <i>Staphylococcus aureus</i>	Change control	BT-CC-1688
Microbiological Performance Criteria/ Entire document	Addition of <i>Staphylococcus aureus</i> ATCC® 25923/ Update to current format and references	Change control	BT-CC-2902