



CERTIFICATE OF ANALYSIS

PRODUCT **MP0255**
BRILLIANCE VRE AGAR

LOT NUMBER 0270

EXPIRY DATE 06.12.2010
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Semi-opaque buff	Conforms to Specification
pH	6.60	6.30 - 6.70
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Enterococcus faecalis ATCC® 29212	Inhibition	Conforms to Specification
Enterococcus faecalis ATCC® 51299	Ppt to 1.5 mm blue colonies	Conforms to Specification
Enterococcus faecium ATCC® 35667	Inhibition	Conforms to Specification
Enterococcus gallinarium ATCC® 49573	Inhibition	Conforms to Specification
Enterococcus casseliflavus ATCC® 25788	Inhibition	Conforms to Specification
Candida albicans ATCC® 10231	Inhibition	Conforms to Specification
Staphylococcus aureus ATCC® 29213	Inhibition	Conforms to Specification
Escherichia coli ATCC® 25922	Inhibition	Conforms to Specification
Klebsiella pneumoniae ATCC® 13883	Inhibition	Conforms to Specification
Citrobacter freundii ATCC® 8090	Inhibition	Conforms to Specification
Enterococcus faecium ATCC® 51559	Ppt to 1.5 mm pink-purple colonies	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager