



CERTIFICATE OF ANALYSIS

PRODUCT PB5003A
ISO SENSITEST AGAR + SHEEP BLOOD
For in vitro diagnostic use.

LOT NUMBER 1027557

EXPIRY DATE 2011.06.10

TYPICAL FORMULATION

	grams per litre
Hydrolysed casein	11.00000
Peptones	3.00000
Glucose	2.00000
Sodium chloride	3.00000
Soluble starch	1.00000
Disodium hydrogen phosphate	2.00000
Sodium acetate	1.00000
Magnesium glycerophosphate	0.20000
Calcium gluconate	0.10000
Cobaltous sulphate	0.00100
Cupric sulphate	0.00100
Zinc sulphate	0.00100
Ferrous sulphate	0.00100
Manganous chloride	0.00200
Menadione	0.00100
Cyanocobalamin	0.00100
L-Cysteine hydrochloride	0.02000
L-Tryptophan	0.02000
Pyridoxine	0.00300
Pantothenate	0.00300
Nicotinamide	0.00300
Biotin	0.00030
Thiamine	0.00004
Adenine	0.01000
Guanine	0.01000
Xanthine	0.01000
Uracil	0.01000
Agar	8.00000
Defibrinated sheep blood	70.00000 ML

General Characteristics	Results	Specification
Colour	Conforms	Carmine red
Appearance	Conforms	Opaque
pH	7.3	7.2 - 7.6
Packaging/Presentation	Conforms	Conforms to specification
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	Within Limits

Susceptibility Testing	Results(mm)	Limits (mm)
S. aureus ATCC@25923 - OX5	38	26 -38



CERTIFICATE OF ANALYSIS

Page 2 of 2

S. aureus ATCC®25923 - E15	34	26 -35
S. aureus ATCC®25923 - VA30	20	16 -23
S. aureus ATCC®25923 - KZ30	40	30 -41
Enterococcus faecalis ATCC®29212 - F100	22	20 -27
Enterococcus faecalis ATCC®29212 - AMP10	34	26 -35
Enterococcus faecalis ATCC®29212 - CIP5	26	20 -27

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature