



CERTIFICATE OF ANALYSIS

PRODUCT PO0148A
MacCONKEY AGAR WITHOUTSALT
1 PACK OF 10 PLATES

LOT NUMBER 1039095

EXPIRY DATE 2008.04.14

Delivery/Customer information
Date Printed 2013.11.11
Delivery No.
Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Dark red
pH (25°C)	7.34	7.20 - 7.60
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Staphylococcus aureus ATCC@25923	50	42	Pale pink cols
Escherichia coli ATCC@25922	83	63	Dark pink cols
Enterococcus faecalis ATCC@29212	30	21	Pink/red cols

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd