



# CERTIFICATE OF ANALYSIS

**PRODUCT**                    **PB0378A**  
**ISO-SEN AGAR + 5%BLD + 20mg/L NAD 4MM**  
**1 PACK OF 10 PLATES**

**LOT NUMBER**                1060946

**EXPIRY DATE**                2008.09.19

Delivery/Customer information	
Date Printed	2013.11.11
Delivery No.	
Customer	
Customer Order number	

Physical Characteristics	Results	Specification
Appearance	Conforms	ruby red
pH (25°C)	7.49	7.20 - 7.60
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Haemophilus influenzae ATCC@49247	42	44	Translucent colonies
Neisseria gonorrhoeae ATCC@49226	68	34	Grey/brown colonies
Streptococcus pneumoniae ATCC@6305	34	37	Grey/green cols, alpha haemolysis
Staphylococcus aureus ATCC@25923	78	51	White/grey cols

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM  
Quality Manager G & M Procter Ltd