



CERTIFICATE OF ANALYSIS

PRODUCT PB0378A
ISO-SEN AGAR + 5%BLD + 20mg/L NAD 4MM
1 PACK OF 10 PLATES

LOT NUMBER 1061280

EXPIRY DATE 2008.09.24

Delivery/Customer information
Date Printed 2013.11.11
Delivery No.
Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	ruby red
pH (25°C)	7.54	7.20 - 7.60
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Haemophilus influenzae ATCC@49247	95	84	Translucent colonies
Neisseria gonorrhoeae ATCC@49226	82	43	Grey/brown colonies
Streptococcus pneumoniae ATCC@6305	48	41	Grey/green cols, alpha haemolysis
Staphylococcus aureus ATCC@25923	54	48	White/grey cols

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd