



CERTIFICATE OF ANALYSIS

PRODUCT PO0121A
CLED AGAR + ANDRADES INDICATOR
1 PACK OF 10 PLATES

LOT NUMBER 1061550

EXPIRY DATE 2008.10.31

Delivery/Customer information
Date Printed 2013.11.11
Delivery No.
Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Blue/green
pH (25°C)	7.58	7.30 - 7.70
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Proteus mirabilis NCTC10975	62	66	Blue/green, translucent cols
Escherichia coli ATCC@25922	71	66	Orange, semi-trans cols, pink halo
Staphylococcus aureus ATCC@25923	87	82	Yellow/orange cols, pink halo

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd