



CERTIFICATE OF ANALYSIS

PRODUCT PO1158A
XLT-4 AGAR
1 PACK OF 10 PLATES

LOT NUMBER 1063073

EXPIRY DATE 2008.11.13

Delivery/Customer information
Date Printed 2013.11.11
Delivery No.
Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Red
pH (25°C)	7.32	7.20 - 7.60
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Salmonella typhimurium ATCC®14028	32	22	red colonies with black centres
Enterococcus faecalis ATCC®29212	>10000	0	No growth
Salmonella enteritidis ATCC®13076	25	24	colonies with black centres

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd