

CERTIFICATE OF ANALYSIS

PRODUCT PB0122A

COLUMBIA HORSE BLOOD AGAR

1 PACK OF 10 PLATES

LOT NUMBER 1064637

EXPIRY DATE 2008.11.10

Delivery/Customer information

Date Printed 2013.11.11 Delivery No.

Customer

Customer Order number

| Physical Characteristics | Results | | Specification |
|--|---------------|------------|----------------------------------|
| Appearance | Conforms | | Red opaque |
| pH (25°C) | 7.5 | | 7.1 - 7.5 |
| Fill Volume/Weight | Conforms | | Conforms to specification |
| Sterility @ 22°C & 32°C ± 2°C for 5 days | Conforms | | Conforms to specification |
| Microbiological Performance | Control c.f.u | Test c.f.u | Specification |
| Streptococcus pyogenes ATCC®19615 | 72 | 62 | Pale straw cols, beta haemolysis |
| Staphylococcus aureus ATCC®25923 | 88 | 80 | cream colonies |

For positive strains, colony count is greater than or equal to 70% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM

Quality Manager G & M Procter Ltd

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology 4 Auld Bond Road, Perth, PH1 3FX