



CERTIFICATE OF ANALYSIS

PRODUCT PB0298A
STREPTOCOCCAL SELECT AGAR C.O.B.A.
1 PACK OF 10 PLATES

LOT NUMBER 1068459

EXPIRY DATE 2008.11.28

Delivery/Customer information

Date Printed
2013.11.11
Delivery No.

Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Red opaque
pH (25°C)	7.4	7.1 - 7.5
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Streptococcus pyogenes ATCC@19615	69	65	Colourless/white cols, β haem
Staphylococcus aureus ATCC@25923	>10000	0	No growth
Streptococcus pneumoniae ATCC@6305	96	56	grey green cols, alpha haemolytic

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd