



CERTIFICATE OF ANALYSIS

PRODUCT PO0299L
CLED AGAR (120MM SQUARE)
1 PACK OF 10 PLATES

LOT NUMBER 1072706

EXPIRY DATE 2009.02.11

Delivery/Customer information
Date Printed 2013.11.11
Delivery No.
Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Blue/green
pH (25°C)	7.4	7.1 - 7.5
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Proteus mirabilis NCTC10975	34	30	Translucent blue cols, no swarming
Escherichia coli ATCC@25922	19	22	Yellow cols
Staphylococcus aureus ATCC@25923	30	25	Yellow cols

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd