



CERTIFICATE OF ANALYSIS

PRODUCT PB0124A
COLUMBIA AGAR + CHOC. HORSE BLOOD
1 PACK OF 10 PLATES

LOT NUMBER 1079041

EXPIRY DATE 2009.04.16

Delivery/Customer information

Date Printed
2013.11.11
Delivery No.

Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Brown opaque
pH (25°C)	7.5	7.1 - 7.5
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Haemophilus influenzae ATCC@49247	62	45	translucent colonies
Neisseria gonorrhoeae ATCC@49226	47	44	Grey/brown colonies
Staphylococcus aureus ATCC@25923	41	40	White/grey colonies

For positive strains, colony count is greater than or equal to 70% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology
4 Auld Bond Road, Perth, PH1 3FX