



# CERTIFICATE OF ANALYSIS

**PRODUCT** PO0140E  
G.C. (VCNT)/SAB.CHLOR BIPLATE  
1 PACK OF 10 PLATES

**LOT NUMBER** 1082418

**EXPIRY DATE** 2009.05.15

### Delivery/Customer information

Date Printed  
2013.11.11  
Delivery No.

Customer  
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Red/Brown translucent
pH (25°C)	7.3	7.1 - 7.5

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Candida albicans ATCC®10231	< 1000	< 1000	Partial inhibition
Proteus hauseri ATCC®13315	>10000	0	No growth
Staphylococcus aureus ATCC®25923	>10000	0	No growth
Neisseria gonorrhoeae ATCC®49226	14	21	Straw/brown-grey colonies

Physical Characteristics	Results	Specification
Appearance bi plate	Conforms	Straw 1-2
pH bi-plate	5.5	5.4 - 5.8

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Candida albicans ATCC10231	65	60	Cream, domed cols
Aspergillus niger ATCC16404	48	51	White mycelia, black spores
Escherichia coli ATCC®25922	>10000	0	No growth

Physical Characteristics	Results	Specification
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

For positive strains, colony count is greater than or equal to 50% of the control medium.



## CERTIFICATE OF ANALYSIS

Page 2 of 2

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

A handwritten signature in black ink, appearing to read "M. E. McCall".

MOIRA MCCALLUM  
Quality Manager G & M Procter Ltd