

CERTIFICATE OF ANALYSIS

PRODUCT CT0127B

CXM30 CEFUROXIME

LOT NUMBER 1084686

EXPIRY DATE 2014.09.30

DATE OF

MANUFACTURE 2011.08.25

Delivery/Customer information

Date Printed 2013.11.11

Delivery No.

Customer

Customer Order number

Microbiological Performance

Assay (mcg/disc) 30.7

Tested in accordance with current CLSI methodology

	Zone Size	Limits
	(mm)	(mm)
Staphylococcus aureus ATCC®25923	35	27 - 35
Escherichia coli ATCC®25922	21	20 - 26
Haemophilus influenzae ATCC®49766	28	28 - 36
Neisseria gonorrhoeae ATCC®49226	33	33 - 41
Tested in accordance with current BSAC methodology		
Escherichia coli ATCC®25922	24	24 - 29
Escherichia coli ATCC®10536	27	25 - 32

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. The results reported were obtained at the time of release.

Lot Accepted. 2011.08.30

Carol MacBride

PRODUCT PERFORMANCE MANAGER

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

