



CERTIFICATE OF ANALYSIS

PRODUCT PO0358A
SABOURAUD DEXTROSE & CHLOR. (DEEP FILL)
1 PACK OF 10 PLATES

LOT NUMBER 1089227

EXPIRY DATE 2009.07.16

Delivery/Customer information
Date Printed 2013.11.11
Delivery No.
Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Straw 1/2
pH (25°C)	5.7	5.4 - 5.8
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Aspergillus brasiliensis ATCC®16404	43	42	White mycelia, black spores
Escherichia coli ATCC®8739	>10000	0	No growth
Candida albicans ATCC®10231	62	49	Cream, domed cols

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd