



CERTIFICATE OF ANALYSIS

PRODUCT PO0659A
R2A AGAR
1 PACK OF 10 PLATES

LOT NUMBER 1097726

EXPIRY DATE 2009.09.30

Delivery/Customer information
Date Printed 2013.11.11
Delivery No.
Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Straw 1
pH (25°C)	7.2	7.0 - 7.4
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Escherichia coli ATCC®8739	74	75	White cols
Bacillus subtilis ATCC®6633	28	27	White cols
Staphylococcus aureus ATCC®6538	59	47	White cols
Pseudomonas aeruginosa ATCC®9027	35	40	White cols

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd