



## CERTIFICATE OF ANALYSIS

**PRODUCT**            **MP0102**  
**ANAEROBIC BRUCELLA AGAR W/LKV**

**LOT NUMBER**            1098

**EXPIRY DATE**            01.07.2011  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear red/brown	Conforms to Specification
pH	7.20	6.80 - 7.20
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

### Microbiological Performance

Organism Tested	Expected Results	Specification
Bacteroides fragilis ATCC® 25285	Growth	Conforms to Specification
Clostridium perfringens ATCC® 13124	Inhibition (partial to complete)	Conforms to Specification
Fusobacterium nucleatum ATCC® 25586	Inhibition (partial to complete)	Conforms to Specification
Escherichia coli ATCC® 25922	Inhibition	Conforms to Specification
Pseudomonas aeruginosa ATCC® 9721	Inhibition (partial to complete)	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Inhibition	Conforms to Specification
Prevotella melaninogenica ATCC® 25845	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell  
Quality Control Manager