



CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT PO0993A
HARLEQUIN SALMONELLA ABC MEDIUM
1 PACK OF 10 PLATES

LOT NUMBER 1104334

EXPIRY DATE 2009.11.09

Delivery/Customer information

Date Printed

2013.11.11

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Straw 1/2
pH (25°C)	7.3	7.0 - 7.4
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Salmonella typhimurium ATCC®14028	41	41	Green cols
Escherichia coli ATCC®25922	< 10000	2	Partial inhibition

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology
4 Auld Bond Road, Perth. PH1 3FX