



G & M Procter Ltd. Certificate of Analysis

PRODUCT PO0779A
ISO-SENSITEST AGAR (4MM DEPTH)
1 PACK OF 10 PLATES

LOT NUMBER 1112291
EXPIRY DATE 2010.02.04
PRODUCTION DATE 2009.12.03
TEST DATE 2009.12.03
REPORTING DATE 2009.12.08

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Straw 2	Straw 2 or straw 3	Appearance and colour
pH (25°C)	7.4	7.2 - 7.6	pH
Fill Volume/Weight	26.1g	25.5 - 26.5g	Fill volume weight check
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Within acceptable limits	Ster. at 22,32,37 & 44°C
The sterility test is performed in accordance with ISO 2859-1: 1999, AQL 1%, general inspection level =1.			

MICROBIOLOGICAL PERFORMANCE

For target organisms the control media must achieve a colony count of 10-100 cfu. The test medium must achieve between 70%-150% of the control medium and show the colonial appearance stated in the specification.

Target Organism	Control c.f.u	Test c.f.u	Colonial Appearance	Colonial Appearance Specification	Accredited Method Reference
Pseudomonas aeruginosa ATCC®27853	33	31	straw colonies	straw colonies	Fertility of Specified Target Organism (s)(Agar)
Enterococcus faecalis ATCC®29212	69	71	Pale straw cols	Pale straw cols	Fertility of Specified Target Organism (s)(Agar)
Staphylococcus aureus ATCC®25923	74	68	Cream cols	Cream cols	Fertility of Specified Target Organism (s)(Agar)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.



Moira McCallum
Quality Manager, G & M Procter Ltd.

2727

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road, Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727



G & M Procter Ltd. Certificate of Analysis

Escherichia coli ATCC®25922	66	62	straw colonies	straw colonies	Fertility of Specified Target Organism (s)(Agar)
--------------------------------	----	----	----------------	----------------	---

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.



Moira McCallum
Quality Manager, G & M Procter Ltd.



CERTIFICATE OF ANALYSIS

Delivery/Customer information

Date Printed

2013.11.11

Delivery No.

Customer

Customer Order number

The information given is believed to be correct. However both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd

Our management system is certified by BSI as being in conformity with ISO 9001:2008, certificate number FM 27644 and ISO 13485:2003, certificate number MD 85850.

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology,
4 Auld Bond Road, Perth, PH1 3FX