



CERTIFICATE OF ANALYSIS

PRODUCT PB0114A
BLOOD AGAR + HORSE BLOOD(BLOOD No 2)
1 PACK OF 10 PLATES

LOT NUMBER 1113900

EXPIRY DATE 2010.01.27

Delivery/Customer information

Date Printed
2013.11.11
Delivery No.

Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Red opaque
pH (25°C)	7.4	7.2 - 7.6
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Streptococcus pyogenes ATCC@19615	77	64	Pale straw cols, beta haemolysis
Staphylococcus aureus ATCC@25923	64	81	White/grey cols
Haemophilus influenzae ATCC@49247	71	73	Pale straw cols

For positive strains, colony count is greater than or equal to 70% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology
4 Auld Bond Road, Perth, PH1 3FX