



# CERTIFICATE OF ANALYSIS

**PRODUCT** PO0696A  
**BURKHOLDERIA CEPACIA MED (MAST)**  
**1 PACK OF 10 PLATES**

**LOT NUMBER** 1125432

**EXPIRY DATE** 2010.05.31

### Delivery/Customer information

Date Printed  
2013.11.11  
Delivery No.

Customer  
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	straw / green
pH (25°C)	6.3	6.0 - 6.4
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Pseudomonas aeruginosa ATCC@27853	>10000	0	Grey cols
Staphylococcus aureus ATCC@25923	>10000	0	No growth
Burkholderia cepacia ATCC@17759	99	98	Sage green cols

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM  
Quality Manager G & M Procter Ltd