



# CERTIFICATE OF ANALYSIS

**PRODUCT** PB0935A  
**STAPH/STREP SELECTIVE AGAR (HPA)**  
**1 PACK OF 10 PLATES**

**LOT NUMBER** 1131203

**EXPIRY DATE** 2010.06.16

### Delivery/Customer information

Date Printed  
2013.11.11  
Delivery No.

Customer  
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Red opaque
pH (25°C)	7.3	7.1 - 7.5
Fill Volume/Weight	Conforms	Conforms to specification
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Staph.aureus var.oxford ATCC@9144	99	95	cream colonies
Streptococcus pyogenes ATCC12344	61	57	Pale straw cols, beta haemolysis
Pseudomonas aeruginosa ATCC@27853	>10000	0	No Growth
Proteus mirabilis NCTC10975	>10000	0	± growth no swarming

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM  
Quality Manager G & M Procter Ltd