



CERTIFICATE OF ANALYSIS

PRODUCT PB0308A
CNA STAPH/STREP SELECT. AGAR (MODIFIED)
1 PACK OF 10 PLATES

LOT NUMBER 1132755

EXPIRY DATE 2010.07.07

Delivery/Customer information
Date Printed 2013.11.11
Delivery No.
Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Red opaque
pH (25°C)	7.4	7.1 - 7.5
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Streptococcus pyogenes ATCC@19615	12	14	Pale straw cols, beta haemolysis
Staphylococcus aureus ATCC@25923	77	63	cream colonies
Proteus mirabilis NCTC10975	>10000	0	± growth no swarming
Pseudomonas aeruginosa ATCC@27853	>10000	0	No Growth

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd