



CERTIFICATE OF ANALYSIS

PRODUCT CT1612B
CPD10 CEFPODOXIME

LOT NUMBER 1138420

EXPIRY DATE 2015.01.31

DATE OF MANUFACTURE 2012.01.05

Delivery/Customer information
Date Printed 2013.11.08
Delivery No.
Customer
Customer Order number

Microbiological Performance

Assay (mcg/disc) 9.95

Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
<i>Staphylococcus aureus</i> ATCC®25923	21	19 - 25
<i>Escherichia coli</i> ATCC®25922	23	23 - 28
<i>Neisseria gonorrhoeae</i> ATCC®49226	43	35 - 43
<i>Haemophilus influenzae</i> ATCC®49247	31	25 - 31
<i>Streptococcus pneumoniae</i> ATCC®49619	34	28 - 34

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. The results reported were obtained at the time of release.

Lot Accepted. 2012.01.11

Carol MacBride
PRODUCT PERFORMANCE MANAGER

ATCC is a registered trade mark of the American Type Culture Collection.
NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



OXOID LIMITED
Wade Road, Basingstoke, Hampshire RG24 8PW. England
oxid@thermofisher.com www.oxid.com