



CERTIFICATE OF ANALYSIS

PRODUCT PO0291A
PSEUDOMONAS C.F.C.SELECTIVE AGAR
1 PACK OF 10 PLATES

LOT NUMBER 1139601

EXPIRY DATE 2010.10.01

Delivery/Customer information
Date Printed 2013.11.11
Delivery No.
Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Straw 1
pH (25°C)	7.2	6.9 - 7.3
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Burkholderia cepacia ATCC@25416	87	80	Straw cols
Staphylococcus aureus ATCC@25923	>10000	0	No growth
Pseudomonas aeruginosa ATCC@27853	32	38	Straw cols, green pigmentation

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd