

CERTIFICATE OF ANALYSIS

	PB1092A	Delivery/Customer information	
PRODUCT		Date Printed	
INCEDEN	ANAEROBE BASAL AGR+NEOMYCIN+AZTREONAM	2013.11.11	
	1 PACK OF 10 PLATES	Delivery No.	
	1140063	Customer	
LOT NUMBER	1140005	Customer Order number	
EXPIRY DATE	2010.09.15		

Physical Characteristics	Results		Specification
Appearance	Conforms		ruby red
pH (25°C)	6.8		6.6 - 7.0
Fill Volume/Weight	Conforms		Conforms to specification
Sterility @ $22^{\circ}C$ & $32^{\circ}C \pm 2^{\circ}C$ for 5 days	Conforms		Conforms to specification
Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Bacteroides fragilis ATCC®25285	17	14	Grey colonies
Escherichia coli ATCC®25922	>10000	0	No growth
Clostridium perfringens ATCC®13124	27	25	Grey colonies
Proteus mirabilis NCTC10975	>10000	0	No growth
Staphylococcus aureus ATCC®25923	>10000	0	No growth

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

M. E.Mcsll

MOIRA MCCALLUM Quality Manager G & M Procter Ltd

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology 4 Auld Bond Road, Perth. PH1 3FX