



CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT PB0135A
GC SELECT. AGAR (VCNT)+ LYSED BLOOD
1 PACK OF 10 PLATES

LOT NUMBER 1144105

EXPIRY DATE 2010.11.09

Delivery/Customer information

Date Printed

2013.11.11

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Red/Brown translucent
pH (25°C)	7.3	7.0 - 7.4
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Candida albicans ATCC®10231	< 1000	< 1000	Partial inhibition
Proteus hauseri ATCC®13315	>10000	0	No growth
Staphylococcus aureus ATCC®25923	>10000	0	No growth
Neisseria gonorrhoeae ATCC®49226	19	18	Grey/brown cols

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology
4 Auld Bond Road, Perth. PH1 3FX