



# G & M Procter Ltd. Certificate of Analysis

**PRODUCT** PB1205A  
**OXOID LYSED GC SELECTIVE AGAR**  
**1 ACK OF 10 PLATES**

**LOT NUMBER** 1148442  
**EXPIRY DATE** 2010.11.25  
**PRODUCTION DATE** 2010.10.14  
**TEST DATE** 2010.10.14  
**REPORTING DATE** 2010.10.19

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Red / Brown translucent	Red / Brown translucent	Appearance and colour
pH (25°C)	7.2	7.0 - 7.4	pH
Fill Volume/Weight	19.6g	18.5 - 20.5g	Fill volume weight check
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Within acceptable limits	Ster. at 22,32,37 & 44°C
The sterility test is performed in accordance with ISO 2859-1: 1999, AQL 1%, general inspection level =1.			

## MICROBIOLOGICAL PERFORMANCE

For target organisms the control media must achieve a colony count of 10-100 cfu. The test medium must achieve between 50%-150% of the control medium and show the colonial appearance stated in the specification. For inhibited organisms, the test medium must show no growth from the stated inoculum.

Target Organism	Control c.f.u	Test c.f.u	Colonial Appearance	Colonial Appearance Specification	Accredited Method Reference
Neisseria gonorrhoeae ATCC®49226	42	35	Grey / brown cols	Grey / brown cols	Fertility of Specified Target Organism (s)(Agar)

Inhibited Organism	Control(cfu)	Test	Specification	Accredited Method Reference
Candida albicans ATCC®10231	10,000 - 100,000	No growth	No growth	Inhibition

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.



Moira McCallum  
Quality Manager, G & M Procter Ltd.

2727

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road, Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727



## G & M Procter Ltd. Certificate of Analysis

Proteus hauseri ATCC®13315	10,000 - 100,000	No growth	No growth	Inhibition
Staphylococcus aureus ATCC®25923	10,000 - 100,000	No growth	No growth	Inhibition

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.



*M. E. McCallum*

Moira McCallum  
Quality Manager, G & M Procter Ltd.