



CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT **PB1065A**
FAA+NAL+TWEEN (HPA)
1 PAK OF 10 PLATES

LOT NUMBER 1157104

EXPIRY DATE 2011.02.24

Delivery/Customer information

Date Printed

2013.11.11

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Red/Brown translucent
pH (25°C)	7.3	7.0 - 7.4
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Bacteroides fragilis ATCC@25285	51	40	Grey colonies
Proteus mirabilis NCTC10975	>10000	0	No growth
Clostridium perfringens ATCC@13124	100000		No growth

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

M. E. McCall

MOIRA MCCALLUM
Quality Manager G & M Procter Ltd

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology
4 Auld Bond Road, Perth. PH1 3FX