



# CERTIFICATE OF ANALYSIS

**PRODUCT** PB1089A  
**ANAEROBE BLOOD AGAR WITH NALIDIXIC ACID**  
**1 PACK OF 10 PLATES**

**LOT NUMBER** 1158959

**EXPIRY DATE** 2011.02.25

Delivery/Customer information
Date Printed 2013.11.11
Delivery No.
Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	ruby red
pH (25°C)	7.3	7.0 - 7.4
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Bacteroides fragilis ATCC@25285	77	75	Grey colonies
Fusobacterium nucleatum NCTC10562	81	78	Grey colonies
Escherichia coli ATCC@25922	>10000	0	No growth

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

MOIRA MCCALLUM  
Quality Manager G & M Procter Ltd