



## CERTIFICATE OF ANALYSIS

**PRODUCT** MT1395  
DERMASEL SLANT 10ML 20 X 125MM

**LOT NUMBER** 1192

**EXPIRY DATE** 31.10.2011  
(dd.mm.yyyy)

### Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Amber slant	Conforms to Specification
pH	6.80	6.70 - 7.10
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance Organism Tested	Expected Results	Specification
A. brasiliensis frm. niger ATCC® 16404	Inhibition (partial to complete)	Conforms to Specification
Escherichia coli ATCC® 25922	Inhibition (partial to complete)	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell  
Quality Control Manager