



G & M Procter Ltd. Certificate of Analysis

PRODUCT PO0779A
ISO-SENSITEST AGAR (4MM DEPTH)
1 PACK OF 10 PLATES

LOT NUMBER 1201495
EXPIRY DATE 2012.03.12
PRODUCTION DATE 2012.01.09
TEST DATE 2012.01.09
REPORTING DATE 2012.01.14

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Straw 2	Straw 2	Appearance and colour
pH (25°C)	7.5	7.2 - 7.6	pH
Fill Volume/Weight	26.0g	25.5 - 26.5g	Fill volume weight check
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Within acceptable limits	Ster. at 22,32,37 & 44°C
The sterility test is performed in accordance with ISO 2859-1: 1999, AQL 1%, general inspection level =1.			

MICROBIOLOGICAL PERFORMANCE

Each organism must achieve a zone size within the range stated in the range specification.

Target Organism	Zone size (mm)	Zone size range (mm)	Accredited Method Reference
Staph aureus Erythromycin E5	30	22 - 31	Zone Diameter (mm)
Staph aureus Gentamicin CN10	27	24 - 30	Zone Diameter (mm)
Staph aureus Penicillin P1	38	32 - 40	Zone Diameter (mm)
Staph aureus Rifampicin RD2	38	27 - 39	Zone Diameter (mm)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.



Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road, Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727



G & M Procter Ltd. Certificate of Analysis

Target Organism	Zone size (mm)	Zone size range (mm)	Accredited Method Reference
Staph aureus Teicoplanin TEC30	22	17 - 23	Zone Diameter (mm)
Staph aureus Vancomycin VA5	17	14 - 20	Zone Diameter (mm)
E coli Ampicillin AMP10	23	21 - 26	Zone Diameter (mm)
E coli Ceftazidime CAZ30	40	32 - 40	Zone Diameter (mm)
E coli Trimethoprim W2.5	32	30 - 37	Zone Diameter (mm)
E coli Cefuroxime CXM30	30	25 - 32	Zone Diameter (mm)
E coli Ciprofloxacin CIP1	39	31 - 40	Zone Diameter (mm)
E coli Gentamicin CN10	27	21 - 27	Zone Diameter (mm)
E faecalis Ampicillin AMP10	34	26 - 35	Zone Diameter (mm)
E faecalis Gentamicin CN200	27	22 - 27	Zone Diameter (mm)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.





G & M Procter Ltd. Certificate of Analysis

Target Organism	Zone size (mm)	Zone size range (mm)	Accredited Method Reference
E faecalis TeicoplaninTEC30	24	19 - 25	Zone Diameter (mm)
E faecalis Vancomycin VA5	15	13 - 19	Zone Diameter (mm)
Pseudomonas Amikacin AK30	30	26 - 32	Zone Diameter (mm)
Pseudomonas Ceftazidime CAZ30	35	27 - 35	Zone Diameter (mm)
Pseudomonas Ciprofloxacin CIP1	27	24 - 30	Zone Diameter (mm)
Pseudomonas Colistin CT25	20	16 - 20	Zone Diameter (mm)
Pseudomonas Gentamicin CN10	25	22 - 28	Zone Diameter (mm)
Pseudomonas Imipenem IPM10	28	23 - 28	Zone Diameter (mm)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.





CERTIFICATE OF ANALYSIS

Delivery/Customer information

Date Printed

2013.11.08

Delivery No.

Customer

Customer Order number

The information given is believed to be correct. However both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

A handwritten signature in black ink that reads "Ian Snowball".

Ian Snowball
Product Performance Manager, G&M Procter Ltd

Our management system is certified by BSI as being in conformity with ISO 9001:2008, certificate number FM 27644 and ISO 13485:2003, certificate number MD 85850.

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology,
4 Auld Bond Road, Perth, PH1 3FX