



CERTIFICATE OF ANALYSIS

PRODUCT **MP0105**
ANAEROBIC BRUCELLA W/5% SHEEP BLOOD

LOT NUMBER 1207

EXPIRY DATE 19.10.2011
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Cherry Red	Conforms to Specification
pH	7.10	6.80 - 7.20
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Bacteroides fragilis ATCC® 25285	Growth	Conforms to Specification
Fusobacterium nucleatum ATCC® 25586	Growth	Conforms to Specification
Clostridium perfringens ATCC® 13124	Growth, beta hemolysis	Conforms to Specification
Peptostreptococcus anaerobius ATCC® 2733	Growth	Conforms to Specification
Prevotella melaninogenica ATCC® 25845	Growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager