



G & M Procter Ltd. Certificate of Analysis

PRODUCT PO1210A
BRILLIANCE MRSA 2 AGAR
1 PACK OF 10 PLATES

LOT NUMBER 1208633
EXPIRY DATE 2012.05.03
PRODUCTION DATE 2012.03.08
TEST DATE 2012.03.09
REPORTING DATE 2012.03.14

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Pale buff	Pale buff	Appearance and colour
pH (25°C)	7.4	7.1 - 7.5	pH
Fill Volume/Weight	18.5g	18.5 - 20.5g	Fill volume weight check
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Within acceptable limits	Ster. at 22,32,37 & 44°C
The sterility test is performed in accordance with ISO 2859-1: 1999, AQL 1%, general inspection level =1.			

MICROBIOLOGICAL PERFORMANCE

For target organisms the control media must achieve a colony count of 10-100 cfu. The test medium must achieve between 50%-150% of the control medium and show the colonial appearance stated in the specification.

For partially inhibited organisms, the test medium must achieve the level of inhibition stated in the specification from the stated inoculum .

For inhibited organisms, the test medium must show no growth from the stated inoculum.

Target Organism	Control c.f.u	Test c.f.u	Colonial Appearance	Colonial Appearance Specification	Accredited Method Reference
Staphylococcus aureus MRSA ATCC®33591	18	21	Blue cols	Blue cols	Fertility of Specified Target Organism (s)(Agar)

Inhibited Organism	Control(cfu)	Test	Specification	Accredited Method Reference
Pseudomonas aeruginosa ATCC®27853	10000	No growth	No growth	Inhibition

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.



Ian Snowball
Ian Snowball
Product Performance Manager
G & M Procter Ltd.

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road,
Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727



G & M Procter Ltd. Certificate of Analysis

Staph epidermidis ATCC®12228	10000	No growth	No growth	Inhibition
------------------------------------	-------	-----------	-----------	------------

Staph.aureus var.oxford ATCC®9144	10000	No growth	No growth	Inhibition
---	-------	-----------	-----------	------------

Proteus mirabilis ATCC®29906	10000	No growth	No growth	Inhibition
------------------------------------	-------	-----------	-----------	------------

Partially Inhibited Organism(s)	Control (cfu)	Test	Specification	Accredited Method Reference
Bacillus licheniformis ATCC®14580	10,000 - 100,000	Pink cols	Pink cols	Partial Inhibition.

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.





CERTIFICATE OF ANALYSIS

Delivery/Customer information

Date Printed

2013.11.08

Delivery No.

Customer

Customer Order number

The information given is believed to be correct. However both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

A handwritten signature in black ink that reads "Ian Snowball".

Ian Snowball
Product Performance Manager, G&M Procter Ltd

Our management system is certified by BSI as being in conformity with ISO 9001:2008, certificate number FM 27644 and ISO 13485:2003, certificate number MD 85850.

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology,
4 Auld Bond Road, Perth, PH1 3FX