



CERTIFICATE OF ANALYSIS

PRODUCT DD0003T
X FACTOR SINGLE

LOT NUMBER 1238274

EXPIRY DATE 2013.09.30

DATE OF MANUFACTURE 2012.09.06

Delivery/Customer information
Date Printed 2013.11.08
Delivery No.
Customer
Customer Order number

Microbiological Performance

Diagnostic characteristics

	Zone Size (mm)	Limits (mm)
<i>Haemophilus influenzae</i> ATCC®19418	10	6 - 13
<i>Haemophilus influenzae</i> ATCC®9334	11	6 - 13
<i>Haemophilus influenzae</i> ATCC®49247	11	6 - 13
<i>Haemophilus parainfluenzae</i> ATCC®33392	0	0 - 0

Quality control ranges are internally derived based on historical data.

Discs are tested with V factor discs.

The half moon zone of growth that appears between the X and V factor discs is measured from the centre of the X disc to the out edge of growth.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. The results reported were obtained at the time of release.

Lot Accepted. 2012.09.07

Carol MacBride
PRODUCT PERFORMANCE MANAGER

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



OXOID LIMITED
 Wade Road, Basingstoke, Hampshire RG24 8PW. England
 oxoid@thermofisher.com www.oxoid.com