



CERTIFICATE OF ANALYSIS

PRODUCT **MP1935**
 SABOURAUD DEXTROSE W/GENTAMICIN

LOT NUMBER 1241

EXPIRY DATE 21.11.2011
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

| Physical Characteristics | Results | Specification |
|------------------------------------|-------------|---------------------------|
| Appearance | Clear amber | Conforms to Specification |
| pH | 5.50 | 5.40 - 5.80 |
| Packaging and Presentation | Conforms | Conforms to specification |
| Sterility @ 35 °C ± 2°C for 3 days | Conforms | Conforms to specification |

Microbiological Performance

| Organism Tested | Expected Results | Specification |
|--|----------------------------------|---------------------------|
| A. brasiliensis frm. niger ATCC® 16404 | Growth | Conforms to Specification |
| Trichophyton mentagrophytes ATCC® 9533 | Growth | Conforms to Specification |
| Escherichia coli ATCC® 25922 | Inhibition (partial to complete) | Conforms to Specification |
| Pseudomonas aeruginosa ATCC® 27853 | Inhibition (partial to complete) | Conforms to Specification |
| Staphylococcus aureus ATCC® 25923 | Inhibition (partial to complete) | Conforms to Specification |
| Candida albicans ATCC® 10231 | Growth | Conforms to Specification |

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager