



# CERTIFICATE OF ANALYSIS

**PRODUCT** CT0058B  
VA30 VANCOMYCIN

**LOT NUMBER** 1241467

**EXPIRY DATE** 2015.09.30

**DATE OF MANUFACTURE** 2012.09.03

Delivery/Customer information
Date Printed 2013.11.08
Delivery No.
Customer
Customer Order number

## Microbiological Performance

Assay (mcg/disc) 32.59

### Tested in accordance with current CLSI methodology

	Zone Size (mm)	Limits (mm)
<i>Staphylococcus aureus</i> ATCC®25923	19	17 - 21
<i>Streptococcus pneumoniae</i> ATCC®49619	22	20 - 27

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. The results reported were obtained at the time of release.

Lot Accepted. 2012.09.05

Carol MacBride  
PRODUCT PERFORMANCE MANAGER

ATCC is a registered trade mark of the American Type Culture Collection.  
NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



**OXOID LIMITED**  
Wade Road, Basingstoke, Hampshire RG24 8PW. England  
oxid@thermofisher.com www.oxid.com