



G & M Procter Ltd. Certificate of Analysis

PRODUCT PO1110A
BRILLIANCE (TM) UTI CLARITY (TM) AGAR
1 PACK OF 10 PLATES

LOT NUMBER 1242139
EXPIRY DATE 2013.01.30
PRODUCTION DATE 2012.12.19
TEST DATE 2012.12.19
REPORTING DATE 2012.12.24

Physical Characteristics	Results	Specification	Accredited Method Reference
pH (25°C)	Straw 2	Straw 2	
Fill Volume/Weight	Straw 2	Straw 2	
Sterility @ 22°C & 32°C ± 2°C for 5 days	7.2	6.8 - 7.2	pH
The sterility test is performed in accordance with ISO 2859-1: 1999, AQL 1%, general inspection level =1.	18.6g	18.5 - 20.5g	Fill volume weight check
	Conforms	Within acceptable limits	Ster. at 22,32,37 & 44°C

MICROBIOLOGICAL PERFORMANCE

For target organisms the control media must achieve a colony count of 10-100 cfu. The test medium must achieve between 50%-150% of the control medium and show the colonial appearance stated in the specification.

Target Organism	Control c.f.u	Test c.f.u	Colonial Appearance	Colonial Appearance Specification	Accredited Method Reference
Proteus mirabilis NCTC10975	85	89	Straw / brown colonies	Straw / brown colonies	Fertility of Specified Target Organism (s)(Agar)
Staphylococcus aureus ATCC®25923	57	49	Cream colonies	Cream colonies	Fertility of Specified Target Organism (s)(Agar)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.



Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road, Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727



G & M Procter Ltd. Certificate of Analysis

Target Organism	Control c.f.u	Test c.f.u	Colonial Appearance	Colonial Appearance Specification	Accredited Method Reference
Enterococcus faecalis ATCC®29212	38	35	Blue / green cols	Blue / green cols	Fertility of Specified Target Organism (s)(Agar)
Escherichia coli ATCC®25922	57	50	Pink colonies	Pink colonies	Fertility of Specified Target Organism (s)(Agar)
Enterobacter aerogenes ATCC®13048	34	29	Blue / purple colonies	Blue / purple colonies	Fertility of Specified Target Organism (s)(Agar)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined. Values are not reported on the Certificate of Analysis but details can be provided on request.





CERTIFICATE OF ANALYSIS

Delivery/Customer information

Date Printed

2013.11.08

Delivery No.

Customer

Customer Order number

The information given is believed to be correct. However both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

A handwritten signature in black ink that reads "Ian Snowball".

Ian Snowball
Product Performance Manager, G&M Procter Ltd

Our management system is certified by BSI as being in conformity with ISO 9001:2008, certificate number FM 27644 and ISO 13485:2003, certificate number MD 85850.

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology,
4 Auld Bond Road, Perth, PH1 3FX