



CERTIFICATE OF ANALYSIS

PRODUCT **MP0351**
COLUMBIA BLOOD AGARW/5% SHEEP BLOOD

LOT NUMBER 1270A

EXPIRY DATE 06.12.2011
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Cherry Red	Conforms to Specification
pH	7.40	7.10 - 7.50
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance	Expected Results	Specification
Organism Tested		
Escherichia coli ATCC® 25922	Growth	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Growth	Conforms to Specification
Streptococcus pyogenes ATCC® 19615	Growth, beta hemolysis	Conforms to Specification
Streptococcus pneumoniae ATCC® 6305	Growth, alpha hemolysis	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager