



CERTIFICATE OF ANALYSIS

PRODUCT **MP1312**
 MACCONKEY AGAR #3 W/ CRYSTAL VIOLET

LOT NUMBER 1270A

EXPIRY DATE 20.12.2011
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Clear pink	Conforms to Specification
pH	7.20	6.90 - 7.30
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
Escherichia coli ATCC® 25922	Growth, pink colonies	Conforms to Specification
Proteus mirabilis ATCC® 12453	Growth colorless cols swarming inhibited	Conforms to Specification
Salmonella typhimurium ATCC® 14028	Growth, colourless colonies.	Conforms to Specification
Enterococcus faecalis ATCC® 29212	Inhibition	Conforms to Specification
Staphylococcus aureus ATCC® 25923	Inhibition	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager