



CERTIFICATE OF ANALYSIS

PRODUCT **MT1943**
SELENITE BROTH 10 ML 16x125MM PCK 100

LOT NUMBER 1270A

EXPIRY DATE 29.05.2012
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	clear yellow broth , oran	Conforms to Specification
pH	7.10	6.90 - 7.30
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance	Expected Results	Specification
Organism Tested		
Escherichia coli ATCC® 25922	Inhibition (partial to complete)	Conforms to Specification
Salmonella typhimurium ATCC® 14028	Growth	Conforms to Specification
Shigella sonnei ATCC® 9290	Variable growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager