



CERTIFICATE OF ANALYSIS

PRODUCT PB0311E
F.A.A. NAT/NEO BIP
1 PACK OF 10 PLATES

LOT NUMBER 1271132

EXPIRY DATE 2013.10.10

Delivery/Customer information
Date Printed 2013.11.08
Delivery No.
Customer
Customer Order number

Physical Characteristics	Results	Specification
Appearance	Conforms	Red/Brown translucent
pH (25°C)	7.3	7.0 - 7.4

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Bacteroides fragilis ATCC@25285	71	61	Grey colonies
Escherichia coli ATCC@25922	>10000	0	No growth

Physical Characteristics	Results	Specification
Appearance bi plate	Conforms	ruby red
pH bi-plate	7.3	7.0 - 7.4

Microbiological Performance	Control c.f.u	Test c.f.u	Specification
Clostridium perfringens ATCC13124	93	79	Grey colonies
Bacteroides fragilis ATCC25285	71	61	Grey colonies
Escherichia coli ATCC25922	>10000	0	No growth

Physical Characteristics	Results	Specification
Fill Volume/Weight	Conforms	Conforms to specification
Sterility @ 22°C & 32°C ± 2°C for 5 days	Conforms	Conforms to specification

For positive strains, colony count is greater than or equal to 50% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.

Ian Snowball
Product Performance Manager, G&M Procter Ltd

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology
4 Auld Bond Road, Perth. PH1 3FX